



DIAGNOSTIC AUTOMATION, INC.

23961 Craftsman Road, Suite D/E/F, Calabasas, CA 91302

Tel: (818) 591-3030 Fax: (818) 591-8383

onestep@rapidtest.com

technicalsupport@rapidtest.com

www.rapidtest.com



See external label



2°C-8°C



Σ=96 tests



Cat # 5114-8

Respiratory syncytial virus IgG ELISA (RSV IgG)

Cat # 5114-8

Enzyme immunoassay based on microtiter plate for the detection and quantitative determination of human IgG antibodies against **RSV** In serum and plasma

Test	RSV IgG ELISA
Method	ELISA: Enzyme Linked Immunosorbent Assay
Principle	Indirect ELISA : Antigen Coated Plate
Detection Range	Qualitative: Positive, Weak Positive, Negative Control
Sample	100µL
Specificity	100%
Sensitivity	86%
Total Time	~110 min
Shelf Life	12-14 months

** Laboratory results can never be the only base of a medical report. The patient history and further tests have to be taken into account.*

Intended Use

The Diagnostic Automation RSV IgG Antibody ELISA Test Kit has been designed for the detection and the quantitative determination of specific IgG antibodies against Respiratory syncytial virus in serum and plasma. Further applications in other body fluids are possible and can be requested from the Technical Service of Diagnostic Automation.

This assay is intended for in-vitro diagnostic use only.

Laboratory results can never be the only base of a medical report. The patient history and further tests have additionally to be taken into account.

General Information

Each year during the winter months, RS viruses spread heavily amongst children and infants. There are recurring infection by RS viruses every year. Voluntary studies with adults have demonstrated that a reinfection with RS viruses is possible. The incubation time is 2 - 6 days. RS viruses are paramyxoviruses with a diameter of 90 - 140 nm. The most noticeable connection of RSV infections with respiratory infections and specific clinical syndromes was detected in infants up to 6 months of age with bronchiolitis or pneumonia. In older infants or small children the disease is milder. In 25 % of infections of the respiratory tract RSV infections are detectable. As reinfections with RSV are possible, it is assumed that these reinfectious antibodies are responsible for the mild course of the disease in adults, being similar to a cold. However, especially in the early years, serum antibodies are no effective protection against infections of the respiratory tract. Therefore, this pathogen may cause bronchiolitis or, in infants up to 4 months of age, pneumonias. Based on their antigen relationship, RSV isolates are differentiated into two major groups (A and B). The surface glycoproteins of the virus (G glycoprotein and fusion glycoprotein) cause the production of virus-neutralizing antibodies. Obviously the G glycoproteins of groups A and B are very different, while the F glycoproteins of both groups show a high antigen concurrence. The complement binding reaction is unsatisfactory for the serological diagnosis of RSV. Enzyme immunoassays are of diagnostic value for the serological diagnosis of RSV infections, as they are very sensitive and allow the differentiation of antigens into the various immunoglobulin classes. In RSV infections, it is possible that the IgM antibody response is missing or so weak that a reliable interpretation of the results is impossible. The detection of IgG antibodies in a single sample is no evidence for an acute infection as, in some patient IgA, antibodies may persist months and years. The method recommended for serological testing of acute RSV infections is the determination of IgG antibodies in serum pairs with significant titer rise.

Principle of the Test

The Diagnostic Automation RSV IgG antibody test kit is based on the principle of the enzyme immunoassay (EIA). RSV antigen is bound on the surface of the microtiter strips. Diluted patient serum or ready-to-use standards are pipetted into the wells of the microtiter plate. A binding between the IgG antibodies of the serum and the immobilized RSV antigen takes place. After a one hour incubation at room temperature, the plate is rinsed with diluted wash solution, in order to remove unbound material. Then ready-to-use anti-human-IgG peroxidase conjugate is added and incubated for 30 minutes. After a further washing step, the substrate (TMB) solution is pipetted and incubated for 20 minutes, inducing the development of a blue dye in the wells. The color development is terminated by the addition of a stop solution, which changes the color from blue to yellow. The resulting dye is measured spectrophotometrically at the wavelength of 450 nm. The concentration of the IgG antibodies is directly proportional to the intensity of the color.

Limitations, Precautions and General Comments

- Only for in-vitro use! Do not ingest or swallow! The usual laboratory safety precautions as well as the prohibition of eating, drinking and smoking in the lab have to be followed.
- All sera and plasma or buffers based upon, have been tested respective to HBsAg, HIV and HCV with recognized methods and were found negative. Nevertheless precautions like the use of latex gloves have to be taken.

- Serum and reagent spills have to be wiped off with a disinfecting solution (e.g. sodium \ hypochlorite, 5%) and have to be disposed of properly.
- All reagents have to be brought to room temperature (18 to 25 °C) before performing the test.
- Before pipetting all reagents should be mixed thoroughly by gentle tilting or swinging. Vigorous shaking with formation of foam should be avoided.
- It is important to pipet with constant intervals, so that all the wells of the microtiter plate have the same conditions.
- When removing reagents out of the bottles, care has to be taken that the stoppers are not contaminated. Further a possible mix-up has to be avoided. The content of the bottles is usually sensitive to oxidation, so that they should be opened only for a short time.
- In order to avoid a carry-over or a cross-contamination, separate disposable pipet tips have to be used.
- No reagents from different kit lots have to be used, they should not be mixed among one another.
- All reagents have to be used within the expiry period.
- In accordance with a Good Laboratory Practice (GLP) or following ISO9001 all laboratory devices employed should be regularly checked regarding the accuracy and precision. This refers amongst others to microliter pipets and washing or reading (ELISA-Reader) instrumentation.
- The contact of certain reagents, above all the stopping solution and the substrate with skin, eye and mucosa has to be avoided, because possible irritations and acid burns could arise, and there exists a danger of intoxication.

Reagents Provided

Store kit components at 2-8°C and do not use after the expiry date on the box outer label. Before use, all components should be allowed to warm up to ambient temperature (18-25°C). After use, the plate should be resealed, the bottle caps replaced and tightened and the kit stored at 2-8°C. The opened kit should be used within three months.

Components	Volume / Qty.
RSV antigen coated microtiter strips	12
Calibrator A (Negative Control)	2 mL
Calibrator B (Cut-Off Standard)	2 mL
Calibrator C (Weak Positive Control)	2 mL
Calibrator D (Positive Control)	2 mL
Enzyme Conjugate	15 mL
Substrate	15 mL
Stop Solution	15 mL
Sample Diluent	60 mL
Washing Buffer (10×)	60 mL
Plastic foils	2
Plastic bag	1

1. Microtiter Strips

12 strips with 8 breakable wells each, coated with a RSV antigen (purified RSV F+G antigen). Ready-to-use.

2. Calibrator A (Negative Control)

2 mL, protein solution diluted with PBS, contains no IgG antibodies against RSV. Addition of 0.01 % methylisothiazolone and 0.01 % bromonitrodioxane. Ready-to-use.

3. Calibrator B (Cut-Off Standard)

2 mL human serum diluted with PBS, contains a low concentration of IgG antibodies against RSV. Addition of 0.01 % methylisothiazolone and 0.01 % bromonitrodioxane. Ready-to-use.

4. Calibrator C (Weak Positive Control)

2 mL, human serum diluted with PBS, contains a medium concentration of IgG antibodies against RSV. Addition of 0.01 % methylisothiazolone and 0.01 % bromonitrodioxane. Ready-to-use.

5. Calibrator D (Positive Control)

2 mL, human serum diluted with PBS, contains a high concentration of IgG antibodies against RSV. Addition of 0.01 % methylisothiazolone and 0.01 % bromonitrodioxane. Ready-to-use.

6. Enzyme Conjugate

15 mL, anti-human-IgG-HRP (rabbit), in protein-containing buffer solution. Addition of 0.01 % methylisothiazolone and 0.01 % bromonitrodioxane and 5 mg/L Proclin™. Ready-to-use.

7. Substrate

15 mL, TMB (tetramethylbenzidine). Ready-to-use.

8. Stop Solution

15 mL, 0.5 M sulfuric acid. Ready-to-use.

9. Sample Diluent

60 mL, PBS/BSA buffer. Addition of 0.095 % sodium azide. Ready-to-use.

10. Washing Buffer

60 mL, PBS + Tween 20, 10x concentrate. Final concentration: dilute 1+9 with distilled water. If during the cold storage crystals precipitate, the concentrate should be warmed up at 37°C for 15 minutes.

11. Plastic Foils

2 pieces to cover the microtiter strips during the incubation.

12. Plastic Bag

Resealable, for the dry storage of non-used strips.

Materials Required but not Provided

- 5 µL-, 100 µL- and 500 µL micro- and multichannel pipets
- Microtiter Plate Reader (450 nm)
- Microtiter Plate Washer
- Reagent tubes for the serum dilution
- Bidistilled water

Specimen Collection and Handling

Principally serum or plasma (EDTA, heparin) can be used for the determination. Serum is separated from the blood, which is aseptically drawn by venipuncture, after clotting and centrifugation. The serum or plasma samples can be stored refrigerated (2-8°C) for up to 48 hours, for a longer storage they should be kept at -20 °C. The samples should not be frozen and thawed repeatedly. Lipemic, hemolytic or bacterially contaminated samples can cause false positive or false negative results.

For the performance of the test the samples (not the standards) have to be diluted 1:101 with ready-to-use sample diluent (e.g. 5 µL serum + 500 µL sample diluent).

Assay Procedure

1. Preparation of Reagents

Washing Solution: dilute before use 1+9 with distilled water. If during the cold storage crystals precipitate, the concentrate should be warmed up at 37°C for 15 minutes.

- Strict adherence to the protocol is advised for reliable performance. Any changes or modifications are the responsibility of the user.
- All reagents and samples must be brought to room temperature before use, but should not be left at this temperature longer than necessary.
- Standards and samples should be assayed in duplicates.
- A standard curve should be established with each assay.
- Return the unused microtiter strips to the plastic bag and store them dry at 2-8°C.

2. Assay Steps

1. Prepare a sufficient amount of microtiter wells for the standards, controls and samples in duplicate as well as for a substrate blank.
2. Pipet 100 µL each of the **diluted** (1:101) samples and the **ready-to-use** standards and controls respectively into the wells. Leave one well empty for the substrate blank.
3. Cover plate with the enclosed foil and incubate at room temperature for 60 minutes.
4. Empty the wells of the plate (dump or aspirate) and add 300 µL of diluted washing solution. This procedure is repeated totally three times. Residuals of the washing buffer are afterwards removed by gentle tapping of the microtiter plate on a tissue cloth.
5. Pipet 100 µL each of ready-to-use conjugate into the wells. Leave one well empty for the substrate blank.
6. Cover plate with the enclosed foil and incubate at room temperature for 30 minutes.
7. Empty the wells of the plate (dump or aspirate) and add 300 µL of diluted washing solution. This procedure is repeated totally three times. Residuals of the washing buffer are afterwards removed by gentle tapping of the microtiter plate on a tissue cloth.
8. Pipet 100 µL each of the ready-to-use substrate into the wells. This time also the substrate blank is pipetted.
9. Cover plate with the enclosed foil and incubate at room temperature for 20 minutes in the dark (e.g. drawer).
10. To terminate the substrate reaction, pipet 100 µL each of the ready-to-use stop solution into the wells. Pipet also the substrate blank.
11. After thorough mixing and wiping the bottom of the plate, perform the reading of the absorption at 450 nm (optionally reference wavelength of 620 nm). The color is stable for at least 60 minutes.

Evaluation

The mean values for the measured absorptions are calculated after subtraction of the substrate blank value. The difference between the single values should not exceed 10%.

Example

	OD Value	corrected OD	Mean OD Value
Substrate Blank	0.015		
Negative Control	0.024 / 0.025	0.009 / 0.010	0.010
Cut-Off Standard	0.460 / 0.481	0.445 / 0.466	0.456
Weak Positive Control	1.084 / 1.050	1.069 / 1.035	1.052
Positive Control	2.213 / 2.169	2.198 / 2.154	2.176

The above table contains only an example, which was achieved under arbitrary temperature and environmental conditions. The described data constitute consequently **no reference values** which have to be found in other laboratories in the same way.

1. Qualitative Evaluation

The calculated absorptions for the patient sera, as mentioned above, are compared with the value for the cut-off standard. If the value of the sample is higher, there is a positive result.

For a value below the cut-off standard, there is a negative result. It seems reasonable to define a range of +/-20 % around the value of the cut-off as a grey zone. In such a case the repetition of the test with the same serum or with a new sample of the same patient, taken after 2-4 weeks, is recommended. Both samples should be measured in parallel in the same run.

The positive control must show at least the double absorption compared with the cut-off standard.

2. Quantitative Evaluation

The ready-to-use standards and controls of the RSV antibody kit are defined and expressed in arbitrary units (U/mL). This results in an exact and reproducible quantitative evaluation. Consequently for a given patient follow-up controls become possible. The values for controls and standards in units are printed on the labels of the vials.

For a quantitative evaluation the absorptions of the standards and controls are graphically drawn against their concentrations. From the resulting reference curve the concentration values for each patient sample can then be extracted in relation to their absorptions. It is also possible to use automatic computer programs.

Assay Characteristics

RSV ELISA	IgG	IgA	IgM
Intra-Assay-Precision	6.4 %	10.1 %	10.1 %
Inter-Assay-Precision	10.7 %	6.8 %	6.8 %
Inter-Lot-Precision	4.8 – 12.1 %	1.0 – 8.7 %	1.0 – 8.7 %
Analytical Sensitivity	1.22 U/mL	1.70 U/mL	1.70 U/mL
Recovery	97 – 110 %	80 – 129 %	89 – 104 %
Linearity	70 – 117 %	87 – 126 %	83 – 121 %
Cross-Reactivity	No cross-reactivity to Influenza A, Parainfluenza 1/2/3 and Adenovirus.		
Interferences	No interferences to bilirubin up to 0.3 mg/mL, hemoglobin up to 8.0 mg/mL and triglycerides up to 5.0 mg/mL		
Clinical Specificity	100 %	100 %	100 %
Clinical Sensitivity	100 %	86 %	100 %

References

1. Grandien M. Paramyxoviridae. The parainfluenza viruses. In: Lennette EH, Halonen P, Murphy FA, editors. Laboratory diagnosis of infectious diseases: principles and practice, vol. 2, Viral, rickettsial, and chlamydial diseases. New York: Springer-Verlag, 1988:484-506.

2. Johnston SLG, Siegel CS. A comparison of direct immunofluorescence, shell vial culture, and conventional cell culture for the rapid detection of Influenza A and B. *Diagn Microbiol Infect Dis* 1991; 14:131-4.
3. Kaul A, et al. Respiratory syncytial virus infection: rapid diagnosis in children by use of indirect immunofluorescence. *Am J Dis Child* 1978; 132:1088-90.
4. Laur BA. Comparison of virus culturing and immunofluorescence for rapid detection of respiratory syncytial virus in nasopharyngeal secretions: sensitivity and specificity. *J Clin Microbiol* 1982; 16:411-2.
5. Palumbo PE, Douglas RG, Jr. Influenza virus. In: Spector S, Lancz G, editors. *Clinical virology manual*. New York: Elsevier, 1986:263-82.
6. Talis A, McIntosh K. Respiratory syncytial virus. In: Balows A, Hausler WJ, Jr, Hermann KL, Isenberg HD, Shadomy HJ, editors. *Manual of clinical microbiology*, 5th ed. Washington, D.C.: American Society for Microbiology, 1991:883-6.
7. Vainionpaa R, Hypia T. Biology of parainfluenza viruses. *Clin Microbiol Rev* 1994; 7:265-75.
8. Waner JL. Parainfluenza virus. In: Balows A, Hausler WJ, Jr., Hermann KL, Isenberg HD, Shadomy HJ, editors. *Manual of clinical microbiology*, 5th ed. Washington D.C.: American Society for Microbiology, 1991:878-82.
9. Wiedbrauk DL, Johnston SLG. Parainfluenza virus. In: *Manual of clinical virology*. New York: Raven Press, 1993.
10. Ehlenfield DR, Cameron K, Welliver RC. Eosinophilia at the time of respiratory syncytial virus bronchiolitis predicts childhood reactive airway disease. *Pediatrics*. 2000; 105(1 Pt 1):79-83.

Date Adopted	Reference No.
2011-02-01	DA-RSV IgG-2011



DIAGNOSTIC AUTOMATION, INC.

23961 Craftsman Road, Suite D/E/F, Calabasas, CA 91302

Tel: (818) 591-3030 Fax: (818) 591-8383

ISO 13485-2003



Revision Date: 04-11-2011