



CORTEZ DIAGNOSTICS, INC.

23961 Craftsman Road, Suite E/F,
 Calabasas, CA 91302 USA
 Tel: (818) 591-3030 Fax: (818) 591-8383
 E-mail: onestep@rapidtest.com
 Web site: www.rapidtest.com

IVD		See external label		2°C-30°C		Σ=25 or 50 tests
REF	Cat. #104116ST-1, 104110ST-6,104110-10					

OneStep RapiCard LH Urine Ovulation Test

For Self Testing

INTRODUCTION

One Step LH Urine Ovulation Test is fast and easy to use. It is a qualitative test that can predict when there is a LH (luteinizing hormone) surge, and in turn, when you are likely to ovulate.

WHEN TO BEGIN TESTING

The first step is to determine the length of your menstrual cycle. This is the number of days from the first day of your menstrual bleeding to the day before your next bleeding starts. Please refer to the chart to determine when you should start testing. If your cycle is shorter than 21 days or longer than 38 days, consult your doctor. If you do not know your cycle length, you may begin testing 11 days after your first period since the average cycle length is 28 days. Perform 1 test each day over a 5 day period, or until the LH surge has been detected.


Your Cycle Length	Start To Test On
21 days	Day 6
22 days	Day 6
23 days	Day 7
24 days	Day 7
25 days	Day 8
26 days	Day 9
27 days	Day 10
28 days	Day 11
29 days	Day 12
30 days	Day 13
31 days	Day 14
32 days	Day 15
33 days	Day 16
34 days	Day 17
35 days	Day 18
36 days	Day 19
37 days	Day 20
38 days	Day 21

URINE COLLECTION

- Do not use first morning urine samples as LH is synthesized in your body early in the morning. It will not show up in your urine until later in the day.
- The best time to collect your urine is between 10am and 8pm.
- Collect urine at about the same time each day.
- Reduce liquid intake about 2 hours before collecting your urine as a diluted urine sample can prevent the test from detecting the LH surge.

TEST PROCEDURE






- Allow the sealed LH cassette test pouch, urine sample, and control solution to reach room temperature (18-30°C).
- Remove the LH Cassette device from the sealed pouch.
- Draw the urine sample to the urine dropper provided.
- Hold the urine dropper upright about 1/2 inch above the sample well of the device. Add 5 drops of urine to this well.
- Lay the test device on a flat surface with the result window facing up. Wait for colored bands to appear. Depending on the concentration of LH in the test specimen, positive results may be observed in as little as 40 seconds. However, to confirm negative results, the complete reaction time of 10 minutes is required. Do not read results after more than 30 minutes.



INTERPRETATION OF RESULTS

- **No LH Surge:** Only one color band appears on the control (C) region or the test (T) band appears but is lighter than the control band. This means there is no LH surge.

- **LH Surge:** Two color bands are visible, and the test (T) band is equal to or darker than the control (C) band. This means that one will probably ovulate in the next 24-48 hours. If trying to get pregnant, the best time to have intercourse is after 24 but before 48 hours.
- **Invalid:** No visible band at all. Repeat with a new test kit.



CORTEZ DIAGNOSTICS, INC.

23961 Craftsman Road, Suite E/F,
Calabasas, CA 91302 USA
Tel: (818) 591-3030 Fax: (818) 591-8383

ISO 13485-2003